

# SUMMARY TEST DATA ON PVVAN-4080-60-MP

Customer: \_\_\_\_\_  
 Job No: \_\_\_\_\_  
 Model No: PVVAN-4080-60-MP  
 Serial No: 13032  
 Part No: \_\_\_\_\_

Tested By: G.Radtke  
 Temperature: +25° C  
 Date: Wednesday, March 17, 2021  
 Drawing No: 27622595 Rev: A1

TEST ITEM NO.	PARAMETERS	SPECIFIED VALUE	TEST RESULTS	QA QC	
1	Frequency Range:	4.0 GHz to 8.0 GHz	4.0 GHz to 8.0 GHz	PMI QA2	
2	Mean Attenuation Range:	60 dB	60.2 dB See Plot		
3	Insertion Loss:	2.5 dB Max.	1.9 dB See Plot		
4	VSWR:	2.0:1 Max	1.4:1 See Plot		
5	Power Rating:	+20 dBm (Operating) +30 dBm (Survival)	Pass		
6	Attenuation Flatness:	± 0.5 dB Typ. @ 10 dB ± 1.0 dB Typ. @ 20 dB ± 1.5 dB Typ. @ 40 dB ± 1.6 dB Typ. @ 60 dB	± 0.24 dB @ 10 dB ± 0.73 dB @ 20 dB ± 1.62 dB @ 40 dB ± 2.12 dB @ 60 dB See Plot		
7	Switching Time:	500 ns Max.	< 80 ns See Typical Characteristics		
8	Control:	10 dB / Volt Analog Controlled	Pass		
9	Power Supply:	+12 VDC @ 100 mA Typ. -12 VDC @ 50 mA Typ.	+12 VDC @ 28 mA -12 VDC @ 0 mA		PMI QA2

Voltage Control (V)	Programmed Attenuation (dB)	Attenuation (dB)	Accuracy of Attenuation (dB)	Flatness (± dB)
1	10	10.25	0.25	0.24
2	20	20.60	0.60	0.73
3	30	30.90	0.90	1.26
4	40	40.98	0.98	1.62
5	50	50.86	0.86	1.87
6	60	60.19	0.19	2.12

QA/QC Approval: \_\_\_\_\_




PMI  
QA2

Date: \_\_\_\_\_

3/17/21

# SUMMARY TEST DATA ON PVVAN-4080-60-MP

13032

